

## **Product Datasheet - Technical Specifications**



More information in our Web-Shop at > www.meilhaus.com

#### Your contact

Technical and commercial sales, price information, quotations, demo/test equipment, consulting:

Tel.:	+49 - (0)81 41 - 52 71-0
FAX:	+49 - (0)81 41 - 52 71-129
E-Mail:	sales@meilhaus.com

Meilhaus Electronic GmbH Tel. Am Sonnenlicht 2 82239 Alling/Germany Mentioned company and product names may be registered trademarks of the respective companies. Errors and omissions excepted. © Meilhaus Electronic.

+49 - (0)81 41 - 52 71-0 Fax +49 - (0)81 41 - 52 71-129 E-Mail sales@meilhaus.com

www.meilhaus.com



# **Baseboard probe**

## P/N: MR09

#### Copyright

#### © 2021, FLIR Systems, Inc.

All rights reserved worldwide. Names and marks appearing herein are either registered trademarks or trademarks of FLIR Systems and/or its subsidiaries. All other trademarks, trade names or company names referenced herein are used for identification only and are the property of their respective owners.

#### **Document identity**

Publ. No.: MR09 Commit: 64574 Language: en-US Modified: 2020-03-18 Formatted: 2021-06-30

#### Website

http://www.flir.com

Customer support

http://support.flir.com

#### Disclaimer

Specifications subject to change without further notice. Camera models and accessories subject to regional market considerations. License procedures may apply. Products described herein may be subject to US Export Regulations. Please refer to exportquestions@flir.com with any questions.



General description		
Part number	MR09	
Part name	Baseboard probe	

When connected to a FLIR Moisture Meter (using the supplied cable), the MR09 allows the user to take relative moisture measurements behind trim pieces, baseboards, crown molding, and masonry; under tile, inside walls/ceilings, and other such places where the sensor blades can be inserted. The blades are designed to 'flex' into the measurement cavity eliminating the need to remove the obstacle that would otherwise prevent a standard pin or pinless measurement. This probe is designed to operate with FLIR Moisture Meters.

General information				
Warranty	https://www.flir.com/testwarranty			
Measurement type	Resistive moisture measurements for reference only			
Penetration depth	15 cm (6 in) maximum			
Meter physical data				
Blade Material	400-series stainless steel; thickness: 1 mm (0.04 in)			
Mounting hole for conversion plate	9.5 mm (3/8 in) threaded			
Cable length	2.45 m (8 ft)			
Weight	494 g (1.2 lb)			
Dimensions ( $H \times W \times L$ )	$34 \times 10 \times 6$ cm (13.4 × 4.0 × 2.4 in.)			
Color	Black handle. Metallic color blades			
Shipping information				
Packaging, type	Retail box			
Packaging, Contents	MR09 Baseboard Probe, RJ45-to-BNC connecting cable, and Instruction sheet			
Each Weight	0.9 kg			
Each Dimensions $(H \times W \times L)$	11.5 × 37.6 × 11.5 cm			
Master Carton Weight	8.5 kg			
Master Carton Dimensions ( $H \times W \times L$ )	$51 \times 25 \times 40$ cm			
Master Carton Quantity	8			
EAN-13	0793950370094			
UPC-12	793950370094			
Country of Origin	China			
Tariff Code	9025.90.0000			



## **Baseboard probe**

P/N: MR09

© 2021, FLIR Systems, Inc. #MR09; r. 64574; en-US

Technical support		
Website	http://support.flir.com	
Accessories		
MR07	Carry Case (optional)	
MR08	Carry Case (optional)	
RJ45-to-BNC Connecting Cable	Included. Connects MR09 to a Moisture meter	

### Compatible with the following products:

- MR160; FLIR MR160: Imaging Moisture Meter with IGM
- MR176; FLIR MR176: Imaging Moisture Meter Plus with IGM
- MR265; FLIR MR265: Moisture Meter and Thermal Imager with MSX
- MR277; FLIR MR277: Building Inspection System with Moisture Hygrometer and MSX IR Camera
- MR60; FLIR MR60: Moisture Meter Pro
- MR77; FLIR MR77: Pinless moisture psychrometer with infrared thermometer and Bluetooth METERLiNK

